




ARC™ STEM



FIRST ARC™ STEM

SIZE 2 STEM USED WITH A 40 MM
HEAD AND 60 MM CUP

NEUTRAL NECK

Dr. Bradley K. Vaughn, M.D., Raleigh, N.C.

Brad does a very conservative posterior approach and has a very experienced surgical crew. They had done one previous intraoperative trial last year with no difficulty.

Dr. Vaughn saw some live surgery last October while visiting in Australia and had been to two cadaver workshops prior to his first case. He had no difficulty at all. He does not take routine x-rays but did on this case. His only comment was the potential need for having an antverted neck available.



IMPLANTING STEM



40 MM C.C. HEAD






TARGET RESTORATION

PHOTOS FROM THIS CASE



No difficulty with exposure with high neck cut in reaming the acetabulum



THE KEY IS THE LEVEL AND ANGLE OF THE NECK OSTEOTOMY. ANGLE 55° LEVEL DOWN FROM SUBCAP 6-8 MM.

SINCE THIS WAS A SIZE 2 STEM THERE WAS A REDUCTION IN SURGICAL TIME EVEN WITH INTRAOPERATIVE X-RAYS.

USED A MUELLER/TSI RASP AFTER AWL THEN WENT TO ARC RASP. ALL WENT VERY WELL. HAS BOOK ADDITIONAL CASES.